



HEARING AID
ACADEMY


2022

BOOK #2 BONUS QUESTIONS




CALL US AT : (903) 487-0097
WWW.HEARINGAIDACADEMY.COM






70) The saturation output refers to :

- A. the limit beyond which the hearing aid will not 'operate properly
- B. the minimum on gain
- C. the full on gain
- D. the maximum gain




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 71

Answer: A




71) Of the following items which would be the most likely to hinder someone from receiving help for their hearing problem:

- A. finances
- B. unaware of their hearing problem
- C. family members
- D. none of the above




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 72

Answer: B




72) Parallel venting to modify the low frequency sounds would be considered:

- A. somewhat important
- B. most types of venting will extend all of the frequencies and make the hearing aid feedback
- C. venting has no effect on the sound of hearing aids
- D. important




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 73

Answer: D




73) If you had to insert tube the ear canal under the headphones this would be for which test:

- A. sound field
- B. free field
- C. for a person with cerumen in the canal when doing pure tones
- D. for a person with a collapsed canal when doing pure tones




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 74

Answer: D




74) If you are trying to find the lowest level at which your patient/client can recognize the signal as speech the test would be called:

- A. SAT speech awareness threshold
- B. SPL speech proficiency level
- C. SRT speech reception threshold
- D. SDT speech discrimination test




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 75

Answer: A




75) A 30 day trial period according to the FDA for hearing aids is:

- A. does not mention
- B. requires
- C. recommended
- D. only would apply for in the home fittings




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 76

Answer: C




76) A good type of question to ask your patient/client to help in their hearing assessment would be:

- A. Do you tend to understand what is said to you if you are facing the speaker?
- B. Do you tend to have problems hearing in groups?
- C. Do you tend to speak loud?
- D. all of the above




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 77

Answer: D




77) Which of the following criteria would be the least likely used when selecting a physician for medical referral:

- A. patient feedback
- B. recommendation of a colleague
- C. trust of physician
- D. location of the physician's office




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 78

Answer: B




78) if your patient has a fenestrated canal you would have to do which of the following when taking ear impressions:

- A. use large or multiple otoblocks
- B. use normal type of otoblock, but with the X-type of string
- C. use normal type of otoblock
- D. none of the above




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 79

Answer: A




79) Some medications causing hearing loss:

- A. would be known as ototoxic drugs
- B. would affect each of your patients differently
- C. would depend on dosage and length of use
- D. would be all of the above




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 80

Answer: D




80) If you would have to worry about the style, material, and the acoustic seal of the earmold selected you may have a patient with:

- A. fenestrated canals
- B. hyperactive movement of the mandibular joint
- C. stenotic canals
- D. atretic canals




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 81

Answer: B




81) If you do a modification to the in-the-ear type of hearing aid you could be changing:

- A. the size of the vent
- B. acoustic feedback
- C. response of the aid
- D. All of the above



Copyright © 2009-2021 – Hearing Aid Academy – All Rights 82


Answer: D



82) If your patient/client is using a device on their hearing aid to exclude room conversation, but is able to hear on the telephone the hearing aid would have an:

- A. MPO switch
- B. MBO switch
- C. MSO switch
- D. MTO switch


Answer: D



83) If you just sold your patient a reconditioned or used hearing aid this would be indicated on the:


- A. contract
- B. package
- C. waiver
- D. all of the above

Answer: A




84) Sudden impact of noise could cause:

- A. acoustic cerebral vascular accident
- B. acoustic neuroma
- C. acoustic trauma
- D. acoustic damping




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 85

Answer: C




85) The FDA regulates:

- A. location of the sell
- B. trial period
- C. cost of the hearing aid
- D. none of the above




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 86

Answer: D




86) An effective verification is:

- A. listening to a watch tick
- B. hearing the words in a song
- C. a measurable improvement in communication
- D. when the patient says he hears better




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 87

Answer: D




87) Functional gain of the instrument:

- A. cannot be calculated
- B. is at or above RTP
- C. is the same as the SSPL90 curve
- D. is one method of fitting verification




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 88

Answer: D




88) Speech tests measured through the audiometer circuit:

- A. cannot be administered because the headset makes the hearing instrument feedback
- B. are 20 dB less because of ANSI Standards
- C. need no conversion when in free field
- D. are 20 dB SPL more because of ANSI Standards




Copyright © 2009-2021 – Hearing Aid Academy – All Rights 89

Answer: C



89) Ideally, when presenting recorded words in noise:

- A. combined words and noise are 70 dBA or less
- B. the noise and words are equal in loudness
- C. the noise is 10 dB louder than the words
- D. the words are 10 dB louder than the noise



Copyright © 2009-2021 – Hearing Aid Academy – All Rights 90

Answer: D



90) Sound field is:

- A. when sound is either absorbed or dissipated before it strikes a reflected surface
- B. a controlled environment
- C. an anechoic chamber
- D. (a) and (c) above

Answer: D